

BIONIME

bionimeusa.com



Technical Data

GM550

| | |
|-------------------------------|---|
| Calibration | Plasma Equivalent |
| Measuring Technology | Glucose oxidase, electrochemical sensor |
| Sample | Capillary whole blood sample |
| Minimum Sample Volume | 1.0 microliter |
| Measuring Range | 10-600 mg/dL |
| Test Time | 5 seconds |
| Memory Capacity | 500 Blood-glucose tests taken with time and date with 7,14, 30 and 90 day averaging |
| Power Saving | Turn off automatically after 2 minutes no use Press the "●" button for 2 seconds |
| Operating Temperature | 10 ~ 40°C (50 ~ 104°F) |
| Coding | No Coding required |
| Operative Relative Humidity | 10 - 90% |
| Hematocrit | 30 - 55% |
| Power Supply | Two CR2032 battery |
| Battery Life | About 1000 tests |
| Meter Dimension | 90.6 mm x 46 mm x 16.5 mm |
| Meter Weight | 53.0 g with batteries |
| Monitor | LCD display |
| Display Area | 33.5 mm x 47.0 mm (WxH) |
| Meter Storage Conditions | -10 ~ 60°C (14 ~ 140 °F) |
| Test Strip Storage Conditions | 4 ~ 30°C (39 ~ 86°F), < 90% relative humidity |

866-326-4401

swissdesign 

© 2009 BIONIME